

User manuals

NucleoMag® DNA Microbiome 744330.1 / .4



Dear valued customer,

Thank you for choosing MACHEREY-NAGEL for your application. We have attached a short protocol for your review.

To obtain the best results we recommend to follow the detailed protocol available online, especially when you are a first time user of this kit.

All important links to the above mentioned product are listed in this leaflet.

Important information regarding product components, specifications, safety instructions, and processing protocols, can all be found on the product website and accessed easily via the QR code.

QR-Code product website



[qr.mn-net.com/qr/\(241\)744330](https://qr.mn-net.com/qr/(241)744330)

Use the following QR code or the link below for direct access to the user manual.

QR-Code user manual



[qr.mn-net.com/qr/\(IFU\)744330](https://qr.mn-net.com/qr/(IFU)744330)

















It is strongly recommended to read the detailed protocol section of the user manual if using the kit for the first time. However, experienced users may refer to the protocol at a glance. The protocol at a glance is designed to be used only as a supplemental tool for quick reference while performing the purification procedure.


We are constantly improving our products and we reserve the right to make changes or additions to protocols. Please check for updated revisions for previously downloaded manuals.

This leaflet does not replace the full manual!



Protocol at a glance

Isolation of DNA from soil, stool and biofilm samples using MN Bead Tubes

1 Prepare sample		MN Bead Tube Type A Up to 200 mg sample material 700 μ L MI1 <i>Optional (Soil): 150 μL Enhancer SX</i> <i>Optional: 2.5 μL Liquid RNase A</i> <i>Optional (Stool/Biofilm): 70 °C, 5 min</i> Agitate on a swing mill or similar device for 1 – 10 min
2 Precipitate debris		Incubate for 5 min at RT 11,000 x g, 5 min Transfer supernatant to a fresh tube
3 Precipitate contaminants		150 μ L MIc Vortex 5 s 10 min, 2–8 °C 11,000 x g, 5 min Transfer supernatant into a fresh tube
4 Bind DNA to NucleoMag® B-Beads	  	Up to 500 μ L lysate 25 μ L NucleoMag® B-Beads 310 μ L MI2 Mix by shaking for 5 min at RT (Optional: Mix by pipetting up and down) Remove supernatant after 5 min separation
5 Wash with MI3	  	Remove Square-well Block from NucleoMag® SEP 600 μ L MI3 Resuspend: Shake 2 min at RT (Optional: Mix by pipetting up and down) Remove supernatant after 2 min separation
6 Wash with MI3	  	Remove Square-well Block from NucleoMag® SEP 600 μ L MI3 Resuspend: Shake 2 min at RT (Optional: Mix by pipetting up and down) Remove supernatant after 2 min separation
7 Wash with MI4	  	Remove Square-well Block on NucleoMag® SEP 600 μ L MI4 Resuspend: Shake 2 min at RT (Optional: Mix by pipetting up and down) Remove supernatant after 2 min separation
8 Wash with 70 % EtOH	  	Remove Square-well Block on NucleoMag® SEP 600 μ L 70 % EtOH Resuspend: Shake 2 min at RT (Optional: Mix by pipetting up and down) Remove supernatant after 2 min separation

9 Dry the beads

10 – 15 min at RT

10 Elute DNARemove Square-well Block from NucleoMag[®]
SEP
50 – 100 μ L MI5Shake 5 min at RT
(Optional: Mix by pipetting up and down)Separate 2 min and transfer DNA into elution
plate

QR-Code Safety Data Sheet

Please find a digital version of the safety data sheets by following the link below:



www.mn-net.com/sds



We strongly recommend to carefully read the detailed protocol section of the product's user manual. If you have any questions about the protocol or product, please contact our Technical Support.

Contact MN

Germany and international

MACHEREY-NAGEL GmbH & Co. KG
Valenciener Str. 11 · 52355 Düren · Germany
Tel.: +49 24 21 969-0
Toll-free: 0800 26 16 000 (Germany only)
E-mail: info@mn-net.com

Technical support Bioanalysis

Tel.: +49 24 21 969-333
E-mail: support@mn-net.com

USA

MACHEREY-NAGEL Inc.
924 Marcon Blvd. · Suite 102 · Allentown PA, 18109 · USA
Toll-free: 888 321 6224 (MACH)
E-mail: sales-us@mn-net.com

France

MACHEREY-NAGEL SAS
1, rue Gutenberg – BP135 · 67720 Hoerdt Cedex · France
Tel.: +33 388 68 22 68
E-mail: sales-fr@mn-net.com

MACHEREY-NAGEL SAS (Société par Actions Simplifiée) au capital de 186600 €

Siret 379 859 531 00020 · RCS Strasbourg B379859531
N° intracommunautaire FR04 379 859 531

Switzerland

MACHEREY-NAGEL AG
Hirsackerstr. 7 · 4702 Oensingen · Switzerland
Tel.: +41 62 388 55 00
E-mail: sales-ch@mn-net.com