

## Customer Information

22.01.2026

# Shipping conditions for BioFix® Lumi Luminous bacteria

All BioFix® Lumi luminous bacteria tests must be stored in a frozen state (see table). After packaging, the tests are shipped on ice packs in an insulated polystyrene box. This shipping packaging ensures a maximum temperature of 0 °C is maintained for four days under normal weather and transport conditions.

Description	REF	Storage temperature
BioFix® Lumi	945023	-21 – -15 °C
BioFix® Lumi	945024	-21 – -15 °C
BioFix® Lumi	945025	-21 – -15 °C
BioFix® Lumi	945002	-21 – -15 °C
BioFix® Lumi	945003	-21 – -15 °C
BioFix® Lumi	945006	-21 – -15 °C
BioFix® Lumi	945007	-21 – -15 °C
BioFix® Lumi Single Shot	945021	-21 – -15 °C
BioFix® Lumi Multi Shot	945022	-21 – -15 °C

Upon arrival, the bacteria must be stored at -21 – -15 °C as quickly as possible. If stored in this manner, including the shipping packaging with ice packs, the goods can be shipped at least once more, provided that the transport time until the next cooling does not exceed four days.

The quality of the luminescent bacteria after transport can be measured using the standard measurements described in DIN EN ISO 11348-2 or DIN EN ISO 11348-3. The inhibition must correspond to the specified tolerance range.

## Contact

If you have further questions, please do not hesitate to contact us:

Technical support and customer service	Tel.:	+49 24 21 969-333
	E-Mail:	support@mn-net.com
For orders (international)	E-Mail:	sales@mn-net.com



Management  
System  
EN ISO 13485:2016  
ISO 9001:2015

[www.tuv.com](http://www.tuv.com)  
ID 0000056401



MACHEREY-NAGEL GmbH & Co. KG · Valenciennes Str. 11 · 52355 Düren · Germany · [www.mn-net.com](http://www.mn-net.com)

DE +49 24 21 969-0 [info@mn-net.com](mailto:info@mn-net.com)  
CH +41 62 388 55 00 [sales-ch@mn-net.com](mailto:sales-ch@mn-net.com)

FR +33 388 68 22 68 [sales-fr@mn-net.com](mailto:sales-fr@mn-net.com)  
US +1 888 321 62 24 [sales-us@mn-net.com](mailto:sales-us@mn-net.com)